

(19) World Intellectual Property Organization  
International Bureau




(43) International Publication Date  
16 May 2002 (16.05.2002)

PCT

(10) International Publication Number  
**WO 02/39106 A1**

- (51) International Patent Classification<sup>7</sup>: **G01N 30/46** (74) Agent: **MARIETTI, Giuseppe**; Marietti, Gislon e Trupiano S.r.l., Via Larga, 16, I-20122 Milan (IT).
- (21) International Application Number: **PCT/IB01/02253**
- (22) International Filing Date:  
28 November 2001 (28.11.2001) (81) Designated States (*national*): JP, US.
- (25) Filing Language: English (84) Designated States (*regional*): European patent (AT, BE, CH, CY, DE, DK, ES, FI, FR, GB, GR, IE, IT, LU, MC, NL, PT, SE, TR).
- (26) Publication Language: English
- (30) Priority Data:  
PCT/IT00/00532  
19 December 2000 (19.12.2000) IT
- (71) Applicant (*for all designated States except US*):  
**THERMO FINNIGAN ITALIA S.p.A.** [IT/IT]; Strada Rivoltana, I-20090 Rodano (IT).
- (72) Inventor; and
- (75) Inventor/Applicant (*for US only*): **BEENS, Jan** [NL/NL]; Eerste Groenelaan 61, NL-1901 TA Castricum (NL).
- Published:**  
— with international search report  
— before the expiration of the time limit for amending the claims and to be republished in the event of receipt of amendments  
— upon request of the applicant, before the expiration of the time limit referred to in Article 21(2)(a)
- For two-letter codes and other abbreviations, refer to the "Guidance Notes on Codes and Abbreviations" appearing at the beginning of each regular issue of the PCT Gazette.*

(54) Title: MODULATOR FOR GAS CHROMATOGRAPHY



(57) **Abstract:** This invention relates to a modulator for use in gas chromatographic analysis, adapted for alternatively trapping and releasing fractions of solutes in a length of a capillary column within a chromatographic oven, characterized in that it comprises at least one nozzle placed to spray at least one jet in at least one corresponding place along said capillary column length, said nozzle (s) being connected each to a source of liquid CO<sub>2</sub> via a related valve, and means for alternatively opening said valve (s) for a predetermined time, to cause a jet of liquid CO<sub>2</sub> to impinge for said predetermined time on said column place and to leave the oven atmosphere to heat said column place after said predetermined time. The modulator can be used in a conventional GC system or in a two dimensional GC system, for modulating the analytes fed to the second capillary column.

**"Modulator for gas chromatography."**

## FIELD OF THE INVENTION

This invention relates to a modulator for modulating sample fractions in a capillary column during a gas chromatographic analysis.

- 5 The modulator according to the present invention can be designed for a traditional gas chromatographic apparatus in order to enhance the sensitivity by narrowing the peaks when placed directly in front of the detector or to focus the injected analytes when placed directly after the injector. However, it can also specially designed for
- 10 comprehensive two dimensional gas chromatography.

## STATE OF THE ART

- The comprehensive two dimensional gas chromatography, also called comprehensive 2D GC, or GCxGC, is a gas chromatographic technique in which the sample is first separated on a conventional
- 15 normal-bore high-resolution capillary GC column in the programmed temperature mode. All of the effluent of this first column is then focused in a large number of extremely narrow (<100 ms) and adjacent fractions at regular, short intervals and subsequently injected onto a second capillary column, which is short and narrow to allow for
- 20 very rapid separations. GCxGC can be interpreted as to exist of two GC systems coupled in series by means of a so-called modulation system (Fig. 1). The first GC is a conventional capillary GC system, including a conventional injector; the second is a fast GC, which is about 50 times faster than the first one. This is accomplished by using a
- 25 short and narrow-bore column to provide very narrow peaks with peak widths at baseline of 100 – 200 ms. The modulation system provides the correspondingly narrow injection pulses in such a way that no sample is lost during the transfer between the chromatographic dimensions. In this way the comprehensive GCxGC technique permits to obtain a

separation power considerably larger than that of conventional capillary gas chromatography, together with an improved sensitivity, a better peak identification and other advantageous features.

As previously said, in order to carry out said GCxGC it is necessary to  
5 operate a so-called modulation system between the first and second capillary column in order to retain and focus the narrow fractions of the effluent of said first column and inject the same at intervals onto said second column.

The most widely used modulators are of the thermal type,  
10 wherein a thermal action on a column length is used to trap and release the fractions to be injected in the second column.

The known heated modulators use an intermediate, thick film modulation capillary to trap (parts of) the eluting analytes from the first column by means of phase-ratio focusing. Heat is applied to thermally  
15 desorb the analytes from the thick film stationary phase in order to re-inject the narrow chemical pulses into the second column. Fig. 2 presents this phase-ratio focusing and thermally desorption process in four steps.

In the first paper describing the comprehensive GCxGC  
20 technique, by Liu and Phillips [Z.Y. Liu, J.B. Phillips, *J. Chrom. Sci.*, **1991**, 29,227-231] and in the Phillips patent[US 5.196.039] a dual-stage metal-coated capillary with a thick film of stationary phase, connected with the outlet of the first column, but placed outside the oven, was employed as a modulation system. Sequentially the two parts of the  
25 metal coated capillary were resistively heated to desorb the analytes trapped due to the lower temperature of the modulation column and its thick stationary phase film. This system appeared not to be robust enough for long use and introduced limitations in the lower temperature of the oven housing of the two columns(as the minimum

temperature of the oven should be in this case at least 100° C higher than the temperature of the modulator which is kept close to the ambient one) .

A more sophisticated heated desorption system was described and made commercially available by Ledford et al. [J.B. Phillips, R.B. Gaines, J. Blomberg, F.W.M. van der Wielen, J.M. Dimandja, V. Green, J. Granger, D. Patterson, L. Racovalis, H.J. de Geus, J. de Boer, P. Haglund, J. Lipsky, V. Sinha, E.B. Ledford, J. High Resolut. Chromatogr., **1999**, 22, 3-10], and Phillips and Ledford patent[US 6.007.602] mainly consisting of a slotted heater moving along the thick film capillary (sweeper) within the gas chromatographic oven.

However, this system too shows drawbacks, mainly due to the movement of the slotted heater in the close vicinity of the tiny capillary, which causes an easy breakage of the column and a limit of the oven maximum temperature.

In order to render more efficient the fraction trapping and eliminate the necessity of a special thick film capillary length, inserted between the first and second column as well as to remove the limitations related with the maximum oven temperature, so called cryogenic or cooled modulators were introduced.

These modulators, consisting of a cold trap moving sequentially forward and backwards along the inlet portion of the second capillary column (the cooling medium sweeps an upstream length of the second column), cryogenically trapping and focusing (parts of) the analytes as they elute from the first column on the first section of the second column itself [R.M. Kinghorn, P.J. Marriott, J. High Resolut. Chromatogr., **1998**, 21, 620-622]. When the cryogenic system moves away from the zone in which the analytes were trapped, the surrounding GC-oven air quickly heats up the trapped analytes

remobilising them for re-injection in the remaining part of the second column. This cryogenic trap, focus and re-injection process is schematically presented in Fig. 3.

The major drawback of this system is the very frequent breakage of the  
5 portion of the fused silica capillary column where the cold trap is moving due to ice formation between the cold trap and the column.

Apart from the mechanical differences between the heated and cooled modulators, there are also some differences in their applicability. In the heated modulators a difference in temperature of  
10 at least 100° C is necessary between the oven and the sweeper, to remobilize the analytes from the thick film capillary that holds the retained fraction. The maximum temperature, to which this capillary can be heated up, i.e. the maximum allowable temperature of its stationary phase, determines the maximum operation temperature of  
15 the sweeper.

The maximum temperature of the column oven will be therefore limited to 100 C below the sweeper temperature and this introduces strong limitations in the application range covered by such systems.

This limitation does not exist with the cooled moving modulator, the  
20 maximum operation temperature of the oven can be much higher as it is limited only by the maximum operating temperature of the two separation columns themselves.

The common characteristic of the thermal modulators as they have been described, however, is the fact that both techniques use a  
25 heating/cooling device that moves across a close distance around a fragile fused silica capillary column. Even very accurate (and rather tedious) tuning of these moving devices and their short distance to the capillaries, frequently leads to breakage of the tiny and fragile capillaries.

Ledford [E.B. Ledford, C. Billesbach, J. High Resol. Chromatogr., 2000, 23, 202-204] introduced a modification of its heating sweeper, by applying a cooling jet of CO<sub>2</sub> on the heating arm. However, this system and the cryogenic system as previously illustrated show all drawbacks of the modulators having movable parts within the oven and moreover the continuous jet of CO<sub>2</sub> tends to create ice formations on the column which involves breaking possibilities and hindering of fraction release.

Ledford (E.B. Ledford, presented on the 23<sup>rd</sup> Symposium on Capillary Gas Chromatography, Riva del Garda, Italy, June 2000) recently proposed a two-stage liquid nitrogen/heated air jet modulator with no moving parts. Two cooling and two heating jets spot-cool and -heat a very short section of the second column to trap/focus and re-inject the modulated fractions. The two cooling jets of the two-stage jet modulator alternately spray liquid nitrogen directly onto the inlet part of the second column for trapping/focusing. Two jets with heated gas alternately heat up these spots to remobilize the analytes for re-injection as very narrow pulses.

The heating jets were necessary, since the temperature of the cooled sections of the second column could reach temperatures as low as -190° C.

Liquid nitrogen is not easily available at every laboratory and needs bulky insulation when transported through tubes. Moreover, the use of liquid nitrogen may create problems due to ice formation within the oven and in particular on the jet nozzles which may such hinder or even stop the release of liquid nitrogen. Moreover, since the hot air jet must have a temperature at least 100° C above the oven temperature and very high air jet temperature cannot be reached for reasons of column integrity (maximum temperature of fused silica columns is 350°

C), this limits the maximum temperature of the oven and the range of applications covered by such systems.

#### OBJECTS OF THE INVENTION

The object of the present invention is now to provide a modulator for  
5 GC or GCxGC which optimises the analytes treatment in a  
conventional GC system and overcomes the drawbacks of the  
presently known modulators for GCxGC, in particular with reference to  
those connected with the mobile modulators (sweepers) and with the  
use of liquid nitrogen and hot air jets in the Ledford modulator with no  
10 moving parts.

#### DESCRIPTION OF THE INVENTION

The main feature and further features of the modulator according to  
this invention are reported in claim 1 and respectively in the  
dependent claims.

#### 15 BRIEF DESCRIPTION OF THE DRAWINGS.

The invention will be more deeply described with reference to the  
accompanying drawings, wherein:

Fig. 1 is a scheme of the GCxGC system.

Fig. 2 is a scheme of the known heating modulation system(sweeper).

20 Fig. 3 is a scheme of the known cryogenic modulation system.

Fig. 4 is a scheme of a modulator according to the present invention.

Fig. 5 is a detail of the jet configuration of the modulator of Fig. 5.

Fig. 6 is a chromatogram obtained by means of a GCxGC separation  
of C<sub>8</sub> through C<sub>18</sub> with a modulator according to the present invention.

25 Fig. 7 is a chromatogram obtained by means of a GCxGC separation  
with a modulator according to the invention and showing the shape  
of the modulated n-C<sub>14</sub> peaks.

Fig. 8 is a scheme of a modulator according to the present invention  
when applied to a conventional GC system.

Fig. 9 represents two chromatograms showing the effect of peak sensitivity enhancement.

Fig. 10a and 10b are diagrammatic representations of an alternative embodiment of the jet configuration, respectively in front view and  
5 side view.

#### DESCRIPTION OF THE PREFERRED EMBODIMENTS

Referring to the drawings, Fig. 1 diagrammatically shows the components of a known GCxGC system, preferably housed in a single oven. Fig. 2 shows a scheme of the heating modulation process, in  
10 which the fraction eluting from the first column is trapped at the upstream end of the thick film of the modulation capillary (phase trapping) (step 1); when the heating sweeper comes in correspondence of this capillary upstream end, the heat effect releases the fraction (step 2) and transports the same along the thick  
15 film capillary, while a further fraction is trapped at the capillary upstream end (step 3).

When the sweeper reaches the second column, the first fraction is released on the same, while the further fraction is still trapped at the modulation capillary upstream end (step 4).

20 Fig. 3 schematically shows the cryogenic modulation process, wherein a cooling medium cools an upstream length of the second column. In correspondence of the cooling medium the fraction is trapped by thermal action and then released when the cooling medium is removed.

25 Fig. 4 is a scheme of a GCxGC system with a modulator according to this invention. The system comprises, in a GC oven 8, an injector 1, a first column 2 and a second column 6 which are connected at 3 according to a well known technique. The second column 6 ends in a detector 7.



On an upstream length 9 of the second column 6 two jets 4A and 4B operate alternatively and at a suitable frequency, which are fed, through corresponding valves 5A and 5B, by a source of liquid CO<sub>2</sub> 10 so that two parts of the capillary length 9 are directly cooled alternating in order to trap and focus the fraction, whereafter they are remobilized by the heat of the surrounding oven air. The opening time of each valve is preferably the same for all valves and half the cycle time, while the opening and closure of the valves are carried out in sequence to cover a cycle time in the order of 0.1 to 30 seconds. It is to be noted that the opening time of said valves could also be different and that this opening time may vary from about 0.1 to about 30 seconds.

The CO<sub>2</sub> jets in Fig.5 consist of two electrical-driven two-way valves 5A, 5B that open and close the liquid-CO<sub>2</sub> line alternating through two pieces of 40 mm long, 0.8 mm ID capillaries 11A/11B, coupled to the nozzles (12A, 12B), 50 mm long 0.5 mm ID capillaries. In order to force as much CO<sub>2</sub> from the outlet of the jets to touch the column, the outlets have been modified to form a slit, 0.04 mm wide and 3 mm long, in parallel above the capillary. To prevent ice formation onto the outside of the jets at oven temperatures below about 100°C, they have been inserted in a 12 mm diameter brass socket to increase the heat capacity.

An alternative embodiment of the jet configuration is shown in figs 10a, 10b and 11, wherein, instead of the slit, the outlet is constructed by inserting a series of seven capillaries in a row between the same brass half blocks. More detailedly, as shown in figs 10a and 10b, each brass block 20 houses a stainless steel capillary 21, for instance having 1/16" OD and 0.7 mm ID, said capillary 21 being connected through a related valve 15, to the CO<sub>2</sub> source 10. Within the end of capillary 21

are inserted for instance seven capillaries 22 placed according to what is shown in fig. 11 and fixed preferably by a ceramic glue or soldering 23, which is able to withstand temperatures of up to 400°C. In the shown example the capillaries have the following dimensions:

- 5 length 35 mm, OD 0.23 mm, ID 0.11 mm and their free portions are aligned so to run in parallel with the secondary GC column 9 so that an optimum heat exchange is enabled by generating a "curtain" of expanding CO<sub>2</sub>.

The axes of the outlet openings of the capillaries 22 are placed 0.4  
10 mm apart, so that the total length of the nozzle again is 3 mm. Of course, the above stated number and dimensions of capillaries can be changed at will.

The above stated construction allows to decrease the consumption of CO<sub>2</sub> and optimize the effectiveness of the throttling process at the  
15 nozzle outlet of the cryogenic jets.

As the liquid CO<sub>2</sub> expands at the outlet of the nozzles, the throttling process cools the departing gas through the Joule-Thompson effect. Since this gas is sprayed directly onto the second column length 9 at the prevailing flow, the column quickly cools down to  
20 about 100° C below the oven temperature. Closing the valve will immediately stop the cooling process and the surrounding air from the stirred oven will heat up the short cooled section of capillary (about 10 mm) momentarily to oven temperature. The time required to heat the capillary column from cryogenic to oven temperature is only 13 ms for  
25 a normal 100 µm column (15 µm polyimide and 80 µm fused silica walls).

The length 9 of the second column in which the modulation takes place, is stretched and secured between two Valco unions 13 mounted on a bracket 14. The stretching is necessary in order to avoid

vibration of the column caused by the rather intense flow of cold CO<sub>2</sub> that is sprayed onto the column. The unions are mounted onto two bands of 1 mm thick, resilient steel in order to compensate for the difference in thermal expansion of the steel bracket and the fused silica column.

A simple timing device that generates the 24 DC voltages for valve switching controls the modulation process. Modulation times shorter than 0.3 seconds can be established.

In order to test the performance of the modulator according to the invention, a gas chromatograph was used with a split/splitless injector and a Flame Ionisation Detector capable to produce a digital signal sampled at 200 Hz rate. The first dimension column 30 m × 0.32 mm ID was coated with methylsilicon polymer, 0.25 microns film thickness. It was coupled through a press-fit connector to the second column 1.5 m × 0.10 mm ID, which was coated with 0.1 µm BPX50 (SGE International, Ringwood, Australia). The flow was set to 1.0 mL/min through a column head pressure of 170 kPa helium. The columns were temperature programmed from 50°C, 4 min isothermal, 2°C/min to 300°C.

The main functions of the modulator are twofold: focusing small fractions from the effluents of the first column into narrow pulses and re-injection of these pulses into the remaining part of the second column. To judge the performance of the modulator, it is sufficient to measure or calculate the bandwidth of the injected pulses. To judge the performance of the dual jet modulator, a series of n-alkanes (C<sub>8</sub> through C<sub>18</sub>, see Fig. 6) was separated. From calculations of the peaks modulated from n-C<sub>14</sub> (see Fig. 7), the peak widths are  $\sigma = 30$  ms, which is better than second dimension peaks previously reported in the literature for known modulation systems (sweeper and

cryomodulators). The injection bandwidth appeared to be  $\sigma < 10$  ms, which is also better than the injection bandwidths of the known sweeper and cryo modulators.

According to what stated above, the jet modulator of this invention is  
5 very simple in construction and easy to install and maintain. Its control is performed by simply switching one, two or more valves, so that no movable part are foreseen within the oven, thus preventing any column breakage due to movement of the previously known movable modulators.

10 Moreover, it has been ascertained that the ability of the modulator according to the invention to focus the trapped first dimension fractions into narrow pulses is superior to that of the modulators known, tested and described in the prior art.

It is to be finally noted that the present modulator, when designed with  
15 one liquid CO<sub>2</sub> jet only, can act as an injection focusing device and/or as a peak narrowing and then a detector sensitivity enhancing device in a conventional one-dimensional GC system. This configuration is depicted in fig. 8, where a capillary column 2 is conventionally housed in an oven and connected with an injector 2 and a detector 3. A jet of  
20 liquid CO<sub>2</sub> issued by a source outside the oven and controlled by a valve, placed outside the oven, can be foreseen to impinge on a column portion respectively directly after the injector (position A) and/or immediately before the injector (position B).

When in position A, the CO<sub>2</sub> jet allows to focus the injected analytes,  
25 while when in position B the jet enhances the sensitivity of the detector by narrowing the peaks.

This is confirmed by the chromatograms of fig. 9, comparing the detector response under the same conditions respectively without sensitivity enhancement (CO<sub>2</sub> jets in position A and B not operative)

and with sensitivity enhancement (CO<sub>2</sub> jet in position A not operative and CO<sub>2</sub> jet in position B operative). A series of low concentration impurities in a main component are shown in the chromatograms of fig. 9, wherein the upper chromatogram shows the main peak together with a series of low concentration impurities in the conventional way, where the lower chromatogram shows how these impurities are collected by means by the single liquid CO<sub>2</sub> jet in position B (at the time of valve on) and released as a series of sharp peaks (at the time of valve off) at increased peak intensities.

### Claims

1. A modulator for use in gas chromatographic analysis, adapted for alternatively trapping and releasing fractions of solutes in a length of a capillary column within a chromatographic oven, characterized in that it comprises at least one nozzle placed to spray at least one jet in at least one corresponding place along said capillary column length, said nozzle(s) being connected each to a source of liquid CO<sub>2</sub> via a related valve, and means for alternatively opening said valve(s) for a predetermined time, to cause a jet of liquid CO<sub>2</sub> to impinge for said predetermined time on said column place and to leave the oven atmosphere to heat said column place after said predetermined time.
2. A modulator according to claim 1, characterized in that said valve(s) is (are) alternatively opened for a predetermined time within a given cycle time and in that said column place is heated by the oven atmosphere during the remaining cycle time.
3. A modulator according to claim 2, for trapping and releasing in sequence fractions of solutes, characterized in that it comprises at least two nozzles placed to spray liquid CO<sub>2</sub> jets in at least two corresponding separated places along said capillary column length, and means for alternatively opening said valves each for a predetermined time in sequence within a given cycle time, to cause each jet of liquid CO<sub>2</sub> to impinge for said predetermined time on the corresponding column place and to leave the oven atmosphere to heat said column place during the remaining cycle time.
4. A modulator according to claim 3, wherein said predetermined time is the same for all valves.
5. A modulator according to claim 3, wherein said predetermined time is different for at least two of said valves.
6. A modulator according to claim 4 or 5, wherein said predetermined

time is ranging from about 0.1 seconds to about 30 seconds.

7. A modulator according to one of claims 4, 5 or 6, wherein said cycle time is the sum of the predetermined times of all valves.

8. A modulator according to one of the claims 2 to 7, wherein said  
5 cycle time is ranging from about 0.1 seconds to about 30 seconds.

9. A modulator according to one of the preceding claims, wherein each said nozzles has an opening in the form of a slit parallel to said capillary length.

10. A modulator according to claim 9, wherein said slit is about 0,04  
10 mm wide and about 3 mm long.

11. A modulator according to one of claims 1 to 8, wherein each said nozzle is formed by a set of capillaries aligned in parallel to said capillary column length.

12. A modulator according to claim 11, wherein the upstream end  
15 of said capillaries open in a common CO<sub>2</sub> feeding duct, to which the capillaries are glued or soldered.

13. A modulator according to claim 12, wherein said capillaries each have an inner diameter of the order of 0.11 mm and each set forms a curtain having a length of about 3 mm.

20 14. A modulator according to one of the preceding claims, wherein said nozzle(s) is(are) inserted in a metal socket.

15. A modulator according to claim 14, wherein said socket is in the form of a brass tube.

25 16. A modulator according to one of the preceding claims, wherein said column length is mounted in stretched conditions.

17. Use of a modulator according to one of the claims 1 to 15 for modulating the solute fractions issued by a first chromatographic column and to be fed to a second chromatographic column in a comprehensive two dimensional gas chromatographic system.

18. Use of a modulator according to one of the claims 1 or 2 and 8 to 16 for modulating the injected fractions immediately downstream the injector in a gas chromatographic system.
19. Use of a modulator according to one of the claims 1 or 2 and 8 to 16, for modulating the eluting fractions from a gas chromatographic column immediately upstream the detector of a gas chromatographic system.



1/7




Figure 1




Figure 2

2/7




Figure 3




Figure 4

3/7




Figure 5




Figure 8

4/7




Figure 6




Figure 7

5/7




Figure 9

6/7




Figure 10a

Figure 10b

7/7




Figure 11

## INTERNATIONAL SEARCH REPORT

Intern. Application No  
PCT/IB 01/02253

A. CLASSIFICATION OF SUBJECT MATTER  
IPC 7 G01N30/46

According to International Patent Classification (IPC) or to both national classification and IPC

## B. FIELDS SEARCHED

Minimum documentation searched (classification system followed by classification symbols)  
IPC 7 G01N

Documentation searched other than minimum documentation to the extent that such documents are included in the fields searched

Electronic data base consulted during the international search (name of data base and, where practical, search terms used)

EPO-Internal

## C. DOCUMENTS CONSIDERED TO BE RELEVANT

Category *	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
A	E.B. LEDFORD JR. ET AL.: "Jet-cooled thermal modulator for comprehensive multidimensional gas chromatography" JOURNAL OF HIGH RESOLUTION CHROMATOGRAPHY., vol. 23, March 2000 (2000-03), pages 202-204, XP000986878 WILEY VCH, WEINHEIM., DE ISSN: 0935-6304 cited in the application page 202, column 1, paragraph 4 -column 2, paragraph 2 --- -/--	1,17-19

☒ Further documents are listed in the continuation of box C.

☒ Patent family members are listed in annex.

## \* Special categories of cited documents:

\*A\* document defining the general state of the art which is not considered to be of particular relevance

\*E\* earlier document but published on or after the international filing date

\*L\* document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified)

\*O\* document referring to an oral disclosure, use, exhibition or other means

\*P\* document published prior to the international filing date but later than the priority date claimed

\*T\* later document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention

\*X\* document of particular relevance; the claimed invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone

\*Y\* document of particular relevance; the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such documents, such combination being obvious to a person skilled in the art.

\*Z\* document member of the same patent family

Date of the actual completion of the international search

1 March 2002

Date of mailing of the international search report

07/03/2002

Name and mailing address of the ISA

European Patent Office, P.B. 5818 Patentlaan 2  
NL - 2280 HV Rijswijk  
Tel. (+31-70) 340-2040, Tx. 31 651 epo nl,  
Fax: (+31-70) 340-3016

Authorized officer

Zinngrebe, U



## INTERNATIONAL SEARCH REPORT

Intern<sup>g</sup> Application No  
PCT/IB 01/02253

C.(Continuation) DOCUMENTS CONSIDERED TO BE RELEVANT		
Category *	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
A	MARRIOTT P ET AL: "Cryogenic solute manipulation in gas chromatography - the longitudinal modulation approach" TRAC, TRENDS IN ANALYTICAL CHEMISTRY, ANALYTICAL CHEMISTRY. CAMBRIDGE, GB, vol. 18, no. 2, February 1999 (1999-02), pages 114-125, XP004155356 ISSN: 0165-9936 page 116	1,17-19
A	WO 98 21574 A (MELBOURNE INST TECH ;MARRIOT PHILIP (AU)) 22 May 1998 (1998-05-22) page 9, paragraph 2 page 18, line 34 -page 19, line 13	1,17-19

# INTERNATIONAL SEARCH REPORT

Information on patent family members

International Application No  
PCT/IB 01/02253

Patent document cited in search report		Publication date	Patent family member(s)	Publication date
WO 9821574	A	22-05-1998	AU 734848 B2	21-06-2001
			AU 4857097 A	03-06-1998
			WO 9821574 A1	22-05-1998
<hr/>				